

VATTURKAR INDUSTRIAL

E-mail: vatsmeter@vsnl.net, Web: www.vatsmeter.com

Technical Specification Of Conductivity/TDS Meters

Sr. No	<u>Descriptions</u>	Conductivity/TDS Indicator Model:CT- 10
1	Display	4 Digit LCD
2	Programming	Yes
3	Out put relay	N.A.
4	Input	From Vats Cell
5	Meter Size	111 X 106 X68 mm
6	Cut out size	92X92 mm
7	Supply Voltage	70 To 270 VAC+/-10 % VAC 50Hz
8	Scale Factor	N.A.
9	Mounting	Field/Panel
10	Cell Cable Length	3 meters from cell to meter
12	Operating Temp.	0-55 ° C
13	Calibration	Through Programming by selecting Cell Constant
14	Main Cord Length	3 meter (3 core)std
15	Ranges	0.1 to 4 µS/cm 10 to 100 µS/cm 50 to 999 µS/cm 1000 to 10000 µS/cm
16	Line Size/Fitting	¾" ABS "T" (To be mounted in Bypass /Drain position)
17	Accuracy	2% Of FSD